


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	Spectra Analysis Services Ltd	
	Issue No: 010 Issue date: 22 November 2016	
	Unit 2 Olympic Business Centre Paycocke Road Basildon Essex SS14 3EX	Contact: Mr Darren Chinnery Tel: +44 (0)1268 534380 Fax: +44 (0)1268 272985 E-Mail: darrenchinnery@spectra-analysis.co.uk Website: www.spectra-analysis.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Unit 2 Olympic Business Centre Paycocke Road Basildon Essex SS14 3EX	Local contact Mr Darren Chinnery Tel: +44 (0)1268 534380 Fax: +44 (0)1268 272985 E-Mail: darrenchinnery@spectra-analysis.co.uk	Health and Hygiene Asbestos (sampling and testing)	A

Site activities performed away from the locations listed above:

Location details		Activity	Location code
Premises (domestic, commercial, industrial)	Contacts as above	Health and Hygiene Asbestos (sampling and testing)	B
Mobile Laboratories	Contacts as above	Health and Hygiene Asbestos (sampling and testing)	C



2621
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Spectra Analysis Services Ltd
Issue No: 010 Issue date: 22 November 2016

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure No 1)	A, B, C
	Fibre counting	HSG 248:February 2005 (Documented In-House Procedure No 1)	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure No 1)	B, C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Procedure No 2, using stereo-microscopy, polarised light microscopy and dispersion staining	A
END			